

EY30 Micronaire Tester

1.1 Product Description

EY30 using Airflow method test the cotton micronaire data, it has a wide measurement range, simple and quick measuring method, the test results directly displayed on the LCD screen, and can also be transferred to the computer via the USB port, and print out the perfect test report.

1.2 Technical Specifications

- Measuring range: 2.0 to 7.5 micronaire
- Measuring accuracy: 0.03 micronaire (standard lab environment)
- Test time: 15 seconds on average
- Sample mass: 5 ± 0.1 g
- Power: AC 100-240V, 50/60Hz
- PC communication: USB port connection
- Dimensions: L520 x W280 x H300mm
- Weight: 10kg
- Warranty: 24 Months



1.3 Professional Technology

1. Precision flow sensor, a wide range of micronaire measurements, reaching 2.0 to 7.5 micronaire, which is the widest range of instruments test micronaire by Airflow method.
2. Fast test, installed the sample, press test key, you can quickly get the result, the instrument with a USB interface, send the result to Pc to analysis of test results, print the standard test report and can also save the test report to EXCEL format.
3. Calibration is convenient, use standard cotton, through the calibration, not standard laboratory environment also can accurately test micronaire value.

1.4 Testing standards

GB1103-2007, ASTM D1448, BS 3181, ISO 2403

1.5 Test Principle

Airflow method testing cotton micronaire and cotton maturity, exert a certain pressure of air in the cotton sample, because of the fineness of the cotton, through the air flow is different, by measuring the air flow also associated with the micronaire test.

1.6 Application

Cotton cultivation, the acquisition of cotton, cotton spinning and cotton inspection.

1.7 Configuration

Standard configuration: Mainbody

Optional accessories: 1.LABTest analysis software.
2.Standard calibration cotton.

2. Test Report

The test results data can send to PC by USB cable and saved as Excel format, concise and practical, to make the perfect test report.

	A	B	C	D	E	F	G	H	I	J
1										
2										
3	Micronaire Test									
4										
5	Test ORG	EYTEST Laboratory				Instrument	EY30			
6	Test time	2012-12-20				Operator	Edwin Zhang			
7	Temperature	20℃				Humidity	65%			
8	Client	Intertek Laboratory								
9	Sample	the cotton for Intertek Laboratory								
10	Remark	The Grade by Iso standard,The leve by GB standard								
11										
12	Test Data									
13	NO.	Test time			Result		Grade		Level	
14	1	10:12			3.55		B		B1	
15	2	10:16			3.52		B		B1	
16	3	10:22			3.53		B		B1	

Contact Us

EYTEST LIMITED

Factory Addr:3/F,A1 Building,B3 Area Fuhai Industry Park,Fuyong Town,Bao'An District,Shenzhen City.China

Contact: Jack Qiu

Email: jack@eytest.com

Mobile:(86)-18675588906

Tel: (86)755-29606518

Fax:(86)755-29606528